

CERTIFICATE OF ANALYSIS

| | |
|-----------------------|---|
| Product: | BabyBio Ni-IDA 1 ml |
| Storage solution: | 20% ethanol |
| Article number: | 45 655 001, 45 655 002, 45 655 003, 45 655 004 |
| Lot number: | 154144 |
| Resin lot number: | 154013 |
| Manufacturing date: | 2020 OCT 06 |
| Product release date: | 2020 OCT 15 |
| Expire date: | 2025-05 |

| | Measured values | Specifications |
|--|------------------------------|-----------------------------------|
| Capacity before charging with Ni ²⁺ | 60 µmol Cu ²⁺ /ml | 50 – 70 µmol Cu ²⁺ /ml |
| Gap control: (5 ml/min, water, 22 °C) | Passed | No gap |
| Flow test: (7 ml/min, water, 22 °C) | Passed | Stable pressure |
| Visual control: | Passed | No visible defects |

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *2020-10-15*

by


Helena Pahlm

Quality Assurance Manager