


CERTIFICATE OF ANALYSIS

Product:	BabyBio Ni-IDA 1 ml	
Storage solution:	20% ethanol	
Article number:	45 655 001, 45 655 002, 45 655 003, 45 655 004	
Lot number:	154144	obsolete due to name change CC-22-10384
Resin lot number:	154013	
Manufacturing date:	2020 OCT 06	
Product release date:	2020 OCT 15	Philip Eriksson
Expire date:	2025-05	2022-08-31


DocuSigned by:

 158C686C7E5E4A0...

	Measured values	Specifications
Capacity before charging with Ni ²⁺	60 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *2020-10-15*

by


 Helena Pahlm
 Quality Assurance Manager