


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio Co-IDA 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 031, 45 655 032, 45 655 033, 45 655 034</b>
Lot number:	<b>154156</b>
Resin lot number:	<b>154004</b>
Manufacturing date:	<b>2020 OCT 13</b>
Product release date:	<b>2020 OCT 23</b>
Expire date:	<b>2025-05</b>

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Co <sup>2+</sup>	60 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *2020-10-23*  
by

  
Helena Pahlm  
Quality Assurance Manager