

## CERTIFICATE OF ANALYSIS

**Product:** **BabyBio Co-IDA 1 ml**

**Storage solution:** **20% ethanol**

**Article number:** **45 655 031, 45 655 032, 45 655 033, 45 655 034**

**Lot number:** **154156** obsolete due to name change  
CC-22-10384

**Resin lot number:** **154004** Philip Eriksson  
2022-08-31

**Manufacturing date:** **2020 OCT 13**

**Product release date:** **2020 OCT 23**

**Expire date:** **2025-05**

DocuSigned by:  
*Philip Eriksson*  
158C686C7E5E4A0...

	Measured values	Specifications
Capacity before charging with Co <sup>2+</sup>	60 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *2020-10-23*  
by

*Helena Pahlm*  
Helena Pahlm  
Quality Assurance Manager