

CERTIFICATE OF ANALYSIS

Product:	BabyBio Zn-NTA 1 ml	
Storage solution:	20% ethanol	
Article number:	45 655 141, 45 655 142, 45 655 143, 45 655 144	
Lot number:	154172	obsolete due to name change CC-22-10384
Resin lot number:	154025	Philip Eriksson 2022-09-02
Manufacturing date:	2020 OCT 15	DocuSigned by: <i>Philip Eriksson</i> 158C686C7E5E4A0...
Product release date:	2020 OCT 23	
Expire date:	2025-05	

	Measured values	Specifications
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu ²⁺)	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *20201023*

by

Helena Pahlm

 Helena Pahlm
 Quality Assurance Manager