

CERTIFICATE OF ANALYSIS


Product:	BabyBio Zn-NTA 1 ml
Storage solution:	20% ethanol
Article number:	45 655 141, 45 655 142, 45 655 143, 45 655 144
Lot number:	154172
Resin lot number:	154025
Manufacturing date:	2020 OCT 15
Product release date:	2020 OCT 23
Expire date:	2025-05

	Measured values	Specifications
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu ²⁺)	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: 20201023

by


Helena Pahlm
Quality Assurance Manager