

## CERTIFICATE OF ANALYSIS

**Product:** BabyBio Zn-IDA 1 ml

**Storage solution:** 20% ethanol

**Article number:** 45 655 041, 45 655 042, 45 655 043, 45 655 044

**Lot number:** 154181 obsolete due to name change  
CC-22-10384

**Resin lot number:** 154014 Philip Eriksson  
2022-08-31

**Manufacturing date:** 2020 OCT 20 DocuSigned by:  
Philip Eriksson  
158C686C7E5E4A0...

**Product release date:** 2020 OCT 23

**Expire date:** 2025-05

	Measured values	Specifications
Capacity before charging with Zn <sup>2+</sup>	60 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: 20201023  
by

  
Helena Pahlm  
Quality Assurance Manager

