


CERTIFICATE OF ANALYSIS

Product:	BabyBio Zn-IDA 5 ml
Storage solution:	20% ethanol
Article number:	45 655 045, 45 655 046, 45 655 047, 45 655 048
Lot number:	154260
Resin lot number:	154014
Manufacturing date:	2020 NOV 19
Product release date:	<i>2020-11-26</i>
Expire date:	2025-05

	Measured values	Specifications
Capacity before charging with Zn ²⁺	60 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued *2020 11 26*
by 
Helena Pahlm
Quality Assurance Manager