
CERTIFICATE OF ANALYSIS

Product: BabyBio Q 1 ml

Storage solution: 20% ethanol

Article number: 45 100 101, 45 100 102, 45 100 103, 45 100 104

Lot number: 154261

Resin lot number: 153921


Manufacturing date: 2020 NOV 23

Product release date: 2020-11-26

Expire date: 2023-11

	Measured values	Specifications
Ionic capacity	203 µmol Cl ⁻ /ml	180-250 µmol Cl ⁻ /ml
Gap control: (5 mL/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 20201126
by 
Helena Pahlm
Quality Assurance Manager