

CERTIFICATE OF ANALYSIS

Product:	BabyBio Co-IDA 5 ml	
Storage solution:	20% ethanol	
Article number:	45 655 035, 45 655 036, 45 655 037, 45 655 038	
Lot number:	154298	
Resin lot number:	154004	
Manufacturing date:	2020 DEC 10	obsolete due to name change CC-22-10384
Product release date:	2021 01 21	Philip Eriksson 2022-08-31
Expire date:	2025-05	DocuSigned by: <i>Philip Eriksson</i> 158C686C7E5E4A0...

	Measured values	Specifications
Capacity before charging with Co ²⁺	60 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C, tested for 10 minutes)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 2021 Jan 21

by



Afifa Trad
Quality Assurance Director

