
CERTIFICATE OF ANALYSIS

Product:	BabyBio DEAE 1 ml
Storage solution:	20% ethanol
Article number:	45 150 101, 45 150 102, 45 150 103, 45 150 104
Lot number:	154328
Resin lot number:	153821
Manufacturing date:	2021 JAN 07
Product release date:	2021 JAN 21
Expire date:	2025-03

	Measured values	Specifications
Ionic capacity:	142 $\mu\text{mol Cl/ml}$	110-160 $\mu\text{mol Cl/ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 2021-04-14

by



Therese Alenvret
QA Consultant