

CERTIFICATE OF ANALYSIS

Product: **BabyBio DEAE 1 ml**

Storage solution: **20% ethanol**

Article number: **45 150 101, 45 150 102, 45 150 103, 45 150 104**

Lot number: **154328** Obsolete due to name change
CC-22-10384

Resin lot number: **153821** Philip Eriksson
2022-08-31

Manufacturing date: **2021 JAN 07**

Product release date: **2021 JAN 21**

Expire date: **2025-03**

DocuSigned by:
Philip Eriksson
158C686C7E5E4A0...

	Measured values	Specifications
Ionic capacity:	142 µmol Cl/ml	110-160 µmol Cl/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 2021-04-14

by



Therese Alenvret
QA Consultant