
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini DEAE 5 mL	Article number(s):	45 150 105 45 150 106 45 150 107 45 150 108
Lot number:	154343		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*: mL	142 µmol Cl ⁻ /mL	110-160 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2021-01-21

Expiry date (year-month): 2025-03

Manufacturing date (year-month): 2021-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-01

By Philip Eriksson

DocuSigned by:



158C886C7E5E4A0...

Quality Assurance
