

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio TREN 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 211, 45 655 212, 45 655 213, 45 655 214</b>
Lot number:	<b>154344</b>
Resin lot number:	<b>153775</b>
Manufacturing date:	<b>2021 JAN 15</b>
Product release date:	<b>2021 JAN 21</b>
Expire date:	<b>2025-02</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	51 µmol Cu <sup>2+</sup> /ml	50 – 60 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 2021-04-14

by



Therese Alenvret  
QA Consultant