
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ A 1 ml
Storage solution:	20% ethanol
Article number:	45 605 101, 45 605 102, 45 605 103, 45 605 104
Lot number:	154355
Resin lot number:	154065
Manufacturing date:	2021 JAN 19
Product release date:	2021 MAR 30
Expire date:	2023-09

	Measured values	Specifications
Dynamic binding capacity for IgG ¹ .	49.2 mg/ml	> 40 mg/ml
Gap control: (4 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued 2021 MAR 30

by



Johanna Tawe

Quality Assurance Manager