
CERTIFICATE OF ANALYSIS

| | |
|-----------------------|---|
| Product: | BabyBio™ affimAb 1 ml |
| Storage solution: | 20% ethanol |
| Article number: | 45 800 101, 45 800 102, 45 800 103, 45 800 104 |
| Lot number: | 154426 |
| Resin lot number: | 153721 |
| Manufacturing date: | 2021 MAR 09 |
| Product release date: | 2021 MAR 18 |
| Expire date: | 2022-12 |


| | Measured values | Specifications |
|---|------------------------|-----------------------|
| Dynamic binding capacity for IgG ¹ . | 43 mg/ml | > 40 mg/ml |
| Gap control: (4 ml/min, water, 22 °C) | Passed | No gap |
| Maximum flow test: (7 ml/min, water, 22 °C) | Passed | Stable pressure |
| Visual control: | Passed | No visible defects |

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 MAR 18**

by



Johanna Tawe
Quality Assurance Manager