


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio™ Dsalt 5 ml</b>	
Storage solution:	<b>20% ethanol</b>	
Article number:	<b>45 360 105, 45 360 106, 45 360 107, 45 360 108, 45 360 109</b>	
Lot number:	<b>154580</b>	obsolete due to name change CC-22-10384
Resin lot number:	<b>154580</b>	Philip Eriksson 2022-08-31
Manufacturing date:	<b>2021 APR 08</b>	 158C686C7E5E4A0...
Product release date:	<b>2021 APR 20</b>	
Expire date:	<b>2026-04</b>	

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	White to-off white
Gap control: (12 ml/min, water, 22 °C)	Passed	No gap
Flow test: (15 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 APR 26**

by



Therese Alenvret  
QA Consultant