

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio™ Dsalt 5 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 360 105, 45 360 106, 45 360 107, 45 360 108, 45 360 109</b>
Lot number:	<b>154580</b>
Resin lot number:	<b>154580</b>
Manufacturing date:	<b>2021 APR 08</b>
Product release date:	<b>2021 APR 20</b>
Expire date:	<b>2026-04</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	White to-off white
Gap control: (12 ml/min, water, 22 °C)	Passed	No gap
Flow test: (15 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 APR 26**

by



Therese Alenvret  
QA Consultant