

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio Ni-NTA 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 101, 45 655 102, 45 655 103, 45 655 104, 45 655 110</b>
Lot number:	<b>154604</b>
Resin lot number:	<b>600064-3029-025</b>
Manufacturing date:	<b>2021 APR 14</b>
Product release date:	<b>2021 MAY 31</b>
Expire date:	<b>2025-05</b>

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	Turquoise to light blue
Metal ion capacity (Tested with Cu <sup>2+</sup> )	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 MAY 31**

by



Susann Sandström  
QA Manager