


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio NTA 1 ml</b>	
Storage solution:	<b>20% ethanol</b>	
Article number:	<b>45 655 111, 45 655 112, 45 655 113, 45 655 114</b>	
Lot number:	<b>154636</b>	obsolete due to name change CC-22-10384
Resin lot number:	<b>6000064-3029-031</b>	Philip Eriksson 2022-09-02
Manufacturing date:	<b>2021 APR 29</b>	 158C686C7E5E4A0...
Product release date:	<b>2021 JUN 15</b>	
Expire date:	<b>2025-09</b>	

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu <sup>2+</sup> )	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 JUN 15**

by



Susann Sandström  
QA Manager