

CERTIFICATE OF ANALYSIS

Product:	BabyBio NTA 1 ml
Storage solution:	20% ethanol
Article number:	45 655 111, 45 655 112, 45 655 113, 45 655 114
Lot number:	154636
Resin lot number:	6000064-3029-031
Manufacturing date:	2021 APR 29
Product release date:	2021 JUN 15
Expire date:	2025-09

	Measured values	Specifications
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu ²⁺)	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 JUN 15**

by



Susann Sandström
QA Manager