

---

## CERTIFICATE OF ANALYSIS

---

|                       |   |
|-----------------------|---|
| <b>Product:</b>       | <b>BabyBio™ A 1 ml</b>                                |
| Storage solution:     | <b>20% ethanol</b>                                    |
| Article number:       | <b>45 605 101, 45 605 102, 45 605 103, 45 605 104</b> |
| Lot number:           | <b>154696</b>   |
| Resin lot number:     | <b>154468</b>   |
| Manufacturing date:   | <b>2021 MAY 21</b>                                    |
| Product release date: | <b>2021 JUN 07</b>                                    |
| Expire date:          | <b>2024-03</b>  |

---

|   | <b>Measured values</b> | <b>Specifications</b> |
|---|------------------------|-----------------------|
| Dynamic binding capacity for IgG <sup>1</sup> . | 55 mg/ml               | > 40 mg/ml            |
| Gap control:<br>(4 ml/min, water, 22 °C)        | Passed                 | No gap                |
| Maximum flow test:<br>(7 ml/min, water, 22 °C)  | Passed                 | Stable pressure       |
| Visual control:                                 | Passed                 | No visible defects    |

---

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 JUN 14**

by



**Anna Boman**  
QA Manager