
CERTIFICATE OF ANALYSIS

Product: BabyBio Q 1 ml

Storage solution: 20% ethanol

Article number: 45 100 101, 45 100 102, 45 100 103, 45 100 104

Lot number: 154753

Resin lot number: 154497

Manufacturing date: 2021 JUN 18

Product release date: 2021 AUG 12

Expire date: 2026-03

	Measured values	Specifications
Ionic capacity	198 µmol Cl/ml	180-250 µmol Cl/ml
Gap control: (5 mL/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 AUG 12**

by



Anna Boman

QA Manager