

CERTIFICATE OF ANALYSIS

Product:	BabyBio S 5 ml
Storage solution:	20% ethanol, 0.2 M sodium acetate
Article number:	45 200 105, 45 200 106, 45 200 107, 45 200 108
Lot number:	154913
Resin lot number:	154212
Manufacturing date:	2021 AUG 05
Product release date:	2021 AUG 27
Expire date:	2023-04

	Measured values	Specifications
Ionic capacity:	193 $\mu\text{mol H}^+/\text{ml}$	180-250 $\mu\text{mol H}^+/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 AUG 27**

by



Anna Boman
QA Manager