


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio S 5 ml</b>	
Storage solution:	<b>20% ethanol, 0.2 M sodium acetate</b>	
Article number:	<b>45 200 105, 45 200 106, 45 200 107, 45 200 108</b>	
Lot number:	<b>154913</b>	obsolete due to name change CC-22-10384
Resin lot number:	<b>154212</b>	
Manufacturing date:	<b>2021 AUG 05</b>	Philip Eriksson 2022-08-31  158C686C7E5E4A0...
Product release date:	<b>2021 AUG 27</b>	
Expire date:	<b>2023-04</b>	

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity:	193 $\mu\text{mol H}^+/\text{ml}$	180-250 $\mu\text{mol H}^+/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 AUG 27**

by

  
\_\_\_\_\_  
Anna Boman  
QA Manager