
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ A 5 ml
Storage solution:	20% ethanol
Article number:	45 605 105, 45 605 106, 45 605 107, 45 605 108
Lot number:	154931
Resin lot number:	154736
Manufacturing date:	2021 AUG 11
Product release date:	2021 AUG 17
Expire date:	2024-06

	Measured values	Specifications
Dynamic binding capacity for IgG ¹ .	54 mg/ml	> 40 mg/ml
Gap control: (15 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (17 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 AUG 17**

by



Susann Sandström
QA Manager