
CERTIFICATE OF ANALYSIS

Product:	BabyBio Q 5 ml
Storage solution:	20% ethanol
Article number:	45 100 105, 45 100 106, 45 100 107, 45 100 108
Lot number:	154961
Resin lot number:	154497
Manufacturing date:	2021 AUG 20
Product release date:	2021 AUG 27
Expire date:	2026-03

	Measured values	Specifications
Ionic capacity:	198 $\mu\text{mol Cl}^-/\text{ml}$	180-250 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 AUG 27**

by



Anna Boman
QA Manager