

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio ACT 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 400 001, 45 400 002, 45 400 003, 45 400 004</b>
Lot number:	<b>154999</b>
Resin lot number:	<b>154345</b>
Manufacturing date:	<b>2021 AUG 31</b>
Product release date:	<b>2021 SEP 15</b>
Expire date:	<b>2024-01</b>

	<b>Measured values</b>	<b>Specifications</b>
Ligand Conc. of reactive groups:	245 $\mu\text{mol/ml}$	> 200 $\mu\text{mol/ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 SEP 15**

by



**Anna Boman**

QA Manager