


CERTIFICATE OF ANALYSIS

Product:	BabyBio ACT 1 ml		
Storage solution:	20% ethanol		
Article number:	45 400 001, 45 400 002, 45 400 003, 45 400 004		
Lot number:	154999	obsolete due to name change: CC-22-10384	
Resin lot number:	154345	Anna Boman	 3C2F4EE14561428...
Manufacturing date:	2021 AUG 31	2022-08-31	
Product release date:	2021 SEP 15		
Expire date:	2024-01		

	Measured values	Specifications
Ligand Conc. of reactive groups:	245 µmol/ml	> 200 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 SEP 15**

by


 Anna Boman
 QA Manager