

CERTIFICATE OF ANALYSIS

Product:	BabyBio ACT 5 ml	
Storage solution:	20% ethanol	
Article number:	45 400 005, 45 400 006, 45 400 007, 45 400 008	
Lot number:	155000	obsolete due to name change: CC-22-10384
Resin lot number:	154345	Anna Boman
Manufacturing date:	2021 AUG 30	2022-08-31
Product release date:	2021 SEP 15	
Expire date:	2024-01	

DocuSigned by:

3C2F4EE14561428...

	Measured values	Specifications
Conc. of reactive groups:	245 µmol/ml	> 200 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 SEP 15**

by



Anna Boman

QA Manager