

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio ACT 5 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 400 005, 45 400 006, 45 400 007, 45 400 008</b>
Lot number:	<b>155000</b>
Resin lot number:	<b>154345</b>
Manufacturing date:	<b>2021 AUG 30</b>
Product release date:	<b>2021 SEP 15</b>
Expire date:	<b>2024-01</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Conc. of reactive groups:	245 µmol/ml	> 200 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 SEP 15**

by



Anna Boman

QA Manager