

CERTIFICATE OF ANALYSIS

Product: **BabyBio IDA His-tag Screening kit 1 ml**

Article number: **45 700 001**

Storage solution: **20% ethanol**

Lot number: **155087**

Resin in the columns	Batch no.	Manufacturing date
BabyBio Ni-IDA 1ml:	154144	2020 OCT 06
BabyBio Cu-IDA 1ml:	154078	2020 SEP 18
BabyBio Co-IDA 1ml:	154156	2020 OCT 13
BabyBio Zn-IDA 1ml:	154181	2020 OCT 20

Product release date: **2021 SEP 29**

Expire date: **2025-05**

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)		50 – 70 µmol/ml
BabyBio Ni-IDA 1 ml	60 µmol/ml	
BabyBio Cu-IDA 1 ml	60 µmol/ml	
BabyBio Co-IDA 1 ml	60 µmol/ml	
BabyBio Zn-IDA 1 ml	60 µmol/ml	
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 SEP 29**

by



Anna Boman

QA Manager