

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio NiMAC 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 311, 45 655 312, 45 655 313, 45 655 314</b>
Lot number:	<b>155109</b>
Resin lot number:	<b>80001-3029-081</b>
Manufacturing date:	<b>2021 OCT 08</b>
Product release date:	<b>2021 NOV 01</b>
Expire date:	<b>2025-04</b>

	<b>Measured values</b>	<b>Specifications</b>
Cu <sup>2+</sup> binding capacity before charging with Ni <sup>2+</sup>	125 µmol/ml	> 60 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 NOV 01**

by



Anna Boman

QA Manager