
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NiMAC 1 mL	Article number(s):	45 655 311 45 655 312 45 655 313 45 655 314
Lot number:	155109		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	125 µmol Cu ²⁺ /mL	>60 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2021-11-01

Expiry date (year-month): 2025-04

Manufacturing date (year-month): 2021-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-02

By Philip Eriksson

DocuSigned by:



158C886C7E5E4A0...

Quality Assurance
