

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio NTA 5 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 115, 45 655 116, 45 655 117, 45 655 118</b>
Lot number:	<b>155168</b>
Resin lot number:	<b>600064-3032-032</b>
Manufacturing date:	<b>2021 OCT 26</b>
Product release date:	<b>2021 NOV 04</b>
Expire date:	<b>2026-09</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	25 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 NOV 04**

by



**Anna Boman**  
QA Manager