


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio NTA 5 ml</b>	
Storage solution:	<b>20% ethanol</b>	
Article number:	<b>45 655 115, 45 655 116, 45 655 117, 45 655 118</b>	
Lot number:	<b>155168</b>	obsolete due to name change CC-22-10384
Resin lot number:	<b>600064-3032-032</b>	Philip Eriksson 2022-09-02
Manufacturing date:	<b>2021 OCT 26</b>	 158C686C7E5E4A0...
Product release date:	<b>2021 NOV 04</b>	
Expire date:	<b>2026-09</b>	

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	25 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 NOV 04**

by

  
\_\_\_\_\_  
Anna Boman  
QA Manager