
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NTA 5 mL	Article number(s):	45 655 115 45 655 116 45 655 117 45 655 118
Lot number:	155168		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity*:	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2021-10-04

Expiry date (year-month): 2026-09

Manufacturing date (year-month): 2021-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-02

By Philip Eriksson

DocuSigned by:



158C88C7E5E4A0...
Quality Assurance
