

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio Ni-NTA 1 ml</b>
<b>Storage solution:</b>	<b>20% ethanol</b>
<b>Article number:</b>	<b>45 655 101, 45 655 102, 45 655 103, 45 655 104, 45 655 110</b>
<b>Lot number:</b>	<b>155183</b>
<b>Resin lot number:</b>	<b>600064-3029-023</b>
<b>Manufacturing date:</b>	<b>2021 OCT 28</b>
<b>Product release date:</b>	<b>2021 NOV 03</b>
<b>Expire date:</b>	<b>2025-05</b>

---

	<b>Measured values</b>	<b>Specifications</b>
<b>Colour:</b>	Passed	Turquoise to light blue
<b>Metal ion capacity (Tested with Cu<sup>2+</sup>)</b>	25 µmol/ml	20 – 30 µmol/ml
<b>Gap control: (5 ml/min, water, 22 °C)</b>	Passed	No gap
<b>Flow test: (7 ml/min, water, 22 °C)</b>	Passed	Stable pressure
<b>Visual control:</b>	Passed	No visible defects

---

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 NOV 03**

by

  
\_\_\_\_\_  
Anna Boman  
QA Manager