

---

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 SEC Cube** Article number(s): 40 301 050

Lot number: 155293

Storage solution: 20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Thyroglobulin):	0.26	0.20-0.30
Average particle size, Dv50:	55 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2022-01-31

Expiry date (year-month): 2026-12

Manufacturing date (year-month): 2021-12

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-01-31

By Susann Sandström



Quality Assurance

---