

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ Ni-NTA 1 ml	Article number(s):	45 655 101 45 655 102 45 655 103 45 655 104 45 655 110
Lot number:	155320		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Metal binding capacity before charging with Ni <sup>2+*</sup> :	25 µmol Cu <sup>2+</sup> /ml	20-30 µmol Cu <sup>2+</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-01-14

Expiry date (year-month): 2025-09


Manufacturing date (year-month): 2022-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-01-14

By Susann Sandström



Quality Assurance

---