
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ Ni-IDA 1 ml	Article number(s):	45 655 001 45 655 002 45 655 003 45 655 004
Lot number:	155355		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ ∗:	60 µmol Cu ²⁺ /ml	50-70 µmol Cu ²⁺ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

∗Property measured on resin.

Release date (year-month-day): 2022-01-28

Expiry date (year-month): 2025-05

Manufacturing date (year-month): 2022-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2022-01-28

By Anna Boman



Quality Assurance
