
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 100 Q	10 100 001
	10 100 002
	10 100 005
Lot number: 155360	10 100 010
	10 100 050
	10 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	171 $\mu\text{mol Cl}^-/\text{ml}$	140-200 $\mu\text{mol Cl}^-/\text{ml}$
Average particle size, Dv50:	96 μm	90-110 μm
Flow stability: (linear flow rate in a 25 x 200 mm column)*	>1000 cm/hr	\geq 1000 cm/hr

*Property measured on base matrix.

Release date (year-month-day): 2022-01-28

Expiry date (year-month): 2027-01

Manufacturing date (year-month): 2022-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-01-28

By Susann Sandström



Quality Assurance
