
CERTIFICATE OF ANALYSIS

Product:	WorkBeads™ affimAb	Article number(s):	40 800 001 40 800 002 40 800 010 40 800 050 40 800 060
Lot number:	155377		
Storage solution:			20% ethanol

	Measured Value	Specification
Dynamic binding capacity for IgG:	52 mg/ml	>40 mg/ml
Average particle size, Dv50:	51 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-05-11

Expiry date (year-month): 2025-02

Manufacturing date (year-month): 2022-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-05-11

By Susann Sandström

DocuSigned by:

Susann Sandström

80DB57460900400

Quality Assurance
