

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ Ni-IDA 5 ml	Article number(s):	45 655 005 45 655 006 45 655 007 45 655 008
Lot number:	155446		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Metal binding capacity before charging with Ni <sup>2+*</sup> :	64 µmol Cu <sup>2+</sup> /ml	50-70 µmol Cu <sup>2+</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-03-01

Expiry date (year-month): 2026-07

Manufacturing date (year-month): 2022-02

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-03-01

By Anna Boman

  
\_\_\_\_\_

Quality Assurance

---