
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ A 5 ml	Article number(s):	45 605 105 45 605 106 45 605 107 45 605 108
Lot number:	155457		
Storage solution:			20% ethanol

	Measured Value	Specification
Dynamic binding capacity for IgG*:	54	>40 mg/ml
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-03-02	Expiry date (year-month): 2024-06
	Manufacturing date (year-month): 2022-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-03-02

By Anna Boman



Quality Assurance
