
CERTIFICATE OF ANALYSIS

Product:	WorkBeads™ Macro SEC	Article number(s):	40 370 001 40 370 003 40 370 010 40 370 050 40 370 060
Lot number:	155459		
Storage solution:			20% ethanol

	Measured Value	Specification
K _D (Thyroglobolin):	0,42	0.41-0.51
Average particle size, Dv50:	46 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-03-09

Expiry date (year-month): 2027-02


Manufacturing date (year-month): 2022-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-06-15

By Philip Eriksson

DocuSigned by:



158C886C7E5E4A0...
Quality Assurance
