
CERTIFICATE OF ANALYSIS

Product: GoBio™ Mini IDA His-Tag Screening Kit 5 mL
Article number(s): 45 700 002

Lot number: 155526

Kit consists of: Lot number:
GoBio Mini Ni-IDA 5 mL: 155446
GoBio Mini Cu-IDA 5 mL: 154112
GoBio Mini Co-IDA 5 mL: 154298
GoBio Mini Zn-IDA 5 mL: 154260

Storage solution: 20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with metal ion:		
GoBio Mini Ni-IDA*:	64 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Cu-IDA*:	60 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Co-IDA*:	60 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Zn-IDA*:	60 µmol/mL	50-70 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-03-16

Expiry date (year-month): 2025-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-02

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0
Quality Assurance
