

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ Zn-NTA 5 ml	Article number(s):	45 655 145 45 655 146 45 655 147 45 655 148
Lot number:	155563		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Metal binding capacity before charging with Zn <sup>2+</sup> ∗:	25 µmol Cu <sup>2+</sup> /ml	20-30 µmol Cu <sup>2+</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-03-29

Expiry date (year-month): 2025-05

Manufacturing date (year-month): 2022-03

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-03-29

By Jonatan Svedmyr



---

Quality Assurance

---