
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ Cu-NTA 5 ml	Article number(s):	45 655 125 45 655 126 45 655 127 45 655 128
Lot number:	155566		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Cu ^{2+*} :	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-03-30

Expiry date (year-month): 2026-07


Manufacturing date (year-month): 2022-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2022-03-30

By Susann Sandström

DocuSigned by:



8DD857460900400

Quality Assurance
