
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 DEAE	40 150 001
	40 150 002
Lot number: 155607	40 150 010
	40 150 050
	40 150 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	113 µmol Cl ⁻ /ml	110-160 µmol Cl ⁻ /ml
Average particle size, Dv50:	51 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-04-11

Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2022-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-04-26

By Susann Sandström

DocuSigned by:

Susann Sandström

8DDB57460900400...

Quality Assurance
